

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063594.D
 Acq On : 23 Sep 2020 17:18
 Operator : JC/MD
 Sample : L4116-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS034-200922

Quant Time: Sep 24 12:36:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	206807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	387012	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	391540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	164353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	176577	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.55	113	127296	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	10.06	98	496177	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
62) 4-Bromofluorobenzene	12.38	95	197456	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
Target Compounds						
3) Chloromethane	2.04	50	1965	0.634	ug/l	96
27) cis-1,2-Dichloroethene	6.78	96	7463	2.550	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063594.D
Acq On : 23 Sep 2020 17:18
Operator : JC/MD
Sample : L4116-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS034-200922

Quant Time: Sep 24 12:36:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

