

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063598.D
 Acq On : 23 Sep 2020 18:54
 Operator : JC/MD
 Sample : L4083-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20200916

Manual Integrations
 APPROVED

MMDadoda
 9/24/2020 1:38:04 PM

Quant Time: Sep 24 04:02:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	203332	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	377905	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	386604	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	163324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	174825	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
35) Dibromofluoromethane	7.55	113	124555	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
50) Toluene-d8	10.06	98	492045	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	12.38	95	191433	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.73	101	12409	5.478	ug/l #	80
12) 1,1-Dichloroethene	3.70	96	2234	0.974	ug/l	87
27) cis-1,2-Dichloroethene	6.79	96	1900m	0.660	ug/l	
44) Trichloroethene	8.80	130	481964	172.899	ug/l	98
64) Tetrachloroethene	10.60	164	2104	0.799	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063598.D
Acq On : 23 Sep 2020 18:54
Operator : JC/MD
Sample : L4083-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20200916

Manual Integrations
APPROVED

MMDadoda
9/24/2020 1:38:04 PM

Quant Time: Sep 24 04:02:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

