

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063601.D
 Acq On : 23 Sep 2020 20:06
 Operator : JC/MD
 Sample : L4083-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE135D3-20200917

Quant Time: Sep 24 04:12:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	383606	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	386877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166284	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	177190	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
35) Dibromofluoromethane	7.55	113	125689	47.33	ug/l	0.00
Spiked Amount			Recovery	=	94.66%	
50) Toluene-d8	10.06	98	493800	49.45	ug/l	0.00
Spiked Amount			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.38	95	194288	53.08	ug/l	0.00
Spiked Amount			Recovery	=	106.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	14712	6.444	ug/l	95
12) 1,1-Dichloroethene	3.71	96	1513	0.654	ug/l #	75
27) cis-1,2-Dichloroethene	6.79	96	13102	4.518	ug/l	88
44) Trichloroethene	8.80	130	51713	18.276	ug/l	99
45) 1,2-Dichloropropane	9.09	63	11382	3.582	ug/l	97
64) Tetrachloroethene	10.60	164	43284	16.431	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063601.D
Acq On : 23 Sep 2020 20:06
Operator : JC/MD
Sample : L4083-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE135D3-20200917

Quant Time: Sep 24 04:12:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

