

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063612.D
 Acq On : 24 Sep 2020 1:41
 Operator : JC/MD
 Sample : PB131805TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131805TB

Quant Time: Sep 24 05:30:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	202069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	377315	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	385725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	157326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	167080	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
35) Dibromofluoromethane	7.55	113	121643	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	10.06	98	489595	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	12.38	95	189619	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
Target Compounds						
16) Acetone	3.78	43	11961	10.820	ug/l	96
18) Methyl Acetate	4.28	43	3982	1.419	ug/l #	70
20) Methylene Chloride	4.51	84	13018	4.648	ug/l	98
43) Isopropyl Acetate	8.13	43	80190	11.367	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063612.D
Acq On : 24 Sep 2020 1:41
Operator : JC/MD
Sample : PB131805TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131805TB

Quant Time: Sep 24 05:30:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

