

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063613.D
 Acq On : 24 Sep 2020 2:05
 Operator : JC/MD
 Sample : PB131813TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131813TB

Manual Integrations
 APPROVED

MMDadoda
 9/24/2020 1:38:26 PM

Quant Time: Sep 24 05:32:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	200781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	388182	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	392345	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	163020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	170736	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
35) Dibromofluoromethane	7.55	113	124021	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
50) Toluene-d8	10.06	98	498146	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	12.38	95	195931	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
Target Compounds						
16) Acetone	3.78	43	12365	11.257	ug/l	96
18) Methyl Acetate	4.29	43	4721m	1.693	ug/l	
20) Methylene Chloride	4.51	84	10478	3.765	ug/l	85
43) Isopropyl Acetate	8.13	43	77491	10.677	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063613.D
Acq On : 24 Sep 2020 2:05
Operator : JC/MD
Sample : PB131813TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131813TB

Manual Integrations
APPROVED

MMDadoda
9/24/2020 1:38:26 PM

Quant Time: Sep 24 05:32:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

