

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063614.D
 Acq On : 24 Sep 2020 2:29
 Operator : JC/MD
 Sample : PB131813ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131813ZHE#13

Manual Integrations
 APPROVED

MMDadoda
 9/24/2020 1:38:28 PM

Quant Time: Sep 24 05:34:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	211155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	404811	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	421733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	177968	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	181877	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
35) Dibromofluoromethane	7.55	113	131125	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
50) Toluene-d8	10.06	98	527882	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.38	95	211404	54.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.46%	
Target Compounds						
16) Acetone	3.78	43	11108	9.616	ug/l	100
18) Methyl Acetate	4.29	43	4176m	1.424	ug/l	
20) Methylene Chloride	4.51	84	8936	3.053	ug/l	93
43) Isopropyl Acetate	8.13	43	80327	10.613	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063614.D
Acq On : 24 Sep 2020 2:29
Operator : JC/MD
Sample : PB131813ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131813ZHE#13

Manual Integrations
APPROVED

MMDadoda
9/24/2020 1:38:28 PM

Quant Time: Sep 24 05:34:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

