

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063615.D
 Acq On : 24 Sep 2020 2:53
 Operator : JC/MD
 Sample : PB131813ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131813ZHE#14

Quant Time: Sep 24 05:35:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	399466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	405942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178166	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	7.55	113	127974	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	10.06	98	515412	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	12.38	95	199883	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
Target Compounds						
16) Acetone	3.78	43	12293	11.002	ug/l	97
18) Methyl Acetate	4.29	43	4771	1.682	ug/l #	81
20) Methylene Chloride	4.50	84	9413	3.325	ug/l #	84
43) Isopropyl Acetate	8.13	43	79990	10.710	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063615.D
Acq On : 24 Sep 2020 2:53
Operator : JC/MD
Sample : PB131813ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131813ZHE#14

Quant Time: Sep 24 05:35:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

