

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063585.D
 Acq On : 23 Sep 2020 13:43
 Operator : JC/MD
 Sample : L4104-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 24 03:02:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	215151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	404000	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	403360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	170511	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	182793	49.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.82%	
35) Dibromofluoromethane	7.55	113	132899	47.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.04%	
50) Toluene-d8	10.06	98	511477	48.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.28%	
62) 4-Bromofluorobenzene	12.38	95	202996	52.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.32%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063585.D
Acq On : 23 Sep 2020 13:43
Operator : JC/MD
Sample : L4104-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 24 03:02:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN063585.D

Acq: 23 Sep 2020 13:43

Instrument :

MSVOA_N

ClientSampleId :


TRIP-BLANK

Tot Ion:168 Resp: 215151

Ion Ratio Lower Upper

168 100

99 67.7 53.0 79.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70 7.80



#33

1,2-Dichloroethane-d4

Concen: 49.910 ug/l

RT: 7.99 min Scan# 1943

Delta R.T. -0.00 min

Lab File: VN063585.D


Acq: 23 Sep 2020 13:43

Tgt Ion: 65 Resp: 182793

Ion Ratio Lower Upper

65 100

67 49.6 0.0 104.2



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

7.99

80000


60000


40000

20000

0

Time--> 7.90 8.00 8.10





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2118

Delta R.T. -0.00 min

Lab File: VN063585.D

Acq: 23 Sep 2020 13:43

Instrument :

MSVOA_N

ClientSampleId :

TRIP-BLANK


Tot Ion:114 Resp: 404000

Ion Ratio Lower Upper

114 100

63 23.2 0.0 46.4




88 16.3 0.0 33.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 47.518 ug/l

RT: 7.55 min Scan# 1807

Delta R.T. 0.00 min

Lab File: VN063585.D

Acq: 23 Sep 2020 13:43


Tgt Ion:113 Resp: 132899

Ion Ratio Lower Upper

113 100

111 100.9 82.9 124.3



192 17.9 13.9 20.9




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.636 ug/l

RT: 10.06 min Scan# 2588

Delta R.T. 0.00 min

Lab File: VN063585.D

Acq: 23 Sep 2020 13:43

Instrument :

MSVOA_N

ClientSampleId :





TRIP-BLANK

Tot Ion: 98 Resp: 511477

Ion Ratio Lower Upper

98 100

100 63.2 50.7 76.1



#62

4-Bromofluorobenzene

Concen: 52.656 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. -0.00 min

Lab File: VN063585.D

Acq: 23 Sep 2020 13:43




Tgt Ion: 95 Resp: 202996


Ion Ratio Lower Upper

95 100

174 71.1 0.0 140.8

176 68.3 0.0 137.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. -0.00 min
Lab File: VN063585.D
Acq: 23 Sep 2020 13:43

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Tot Ion:117 Resp: 403360
Ion Ratio Lower Upper
117 100
82 61.6 49.2 73.8
119 31.2 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3601
Delta R.T. -0.00 min
Lab File: VN063585.D
Acq: 23 Sep 2020 13:43

Tgt Ion:152 Resp: 170511
Ion Ratio Lower Upper
152 100
115 64.4 32.2 96.6
150 154.7 0.0 343.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

