

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
 Data File : VN068621.D
 Acq On : 23 Sep 2021 15:37
 Operator : JC/MD
 Sample : VN0923WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0923WBL01

Quant Time: Sep 23 16:16:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	636648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1058643	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	902102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	330787	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	576539	58.234	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	116.460%
35) Dibromofluoromethane	8.024	113	398870	56.395	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.780%
50) Toluene-d8	10.443	98	1579218	52.599	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.200%
62) 4-Bromofluorobenzene	12.733	95	469497	45.723	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
Data File : VN068621.D
Acq On : 23 Sep 2021 15:37
Operator : JC/MD
Sample : VN0923WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0923WBL01

Quant Time: Sep 23 16:16:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

