

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
 Data File : VN068623.D
 Acq On : 23 Sep 2021 16:35
 Operator : JC/MD
 Sample : M3826-05 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRIP-COMP-AUD-1181-1187

Quant Time: Sep 23 17:25:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	604581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1089559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	922167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	334703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	542516	57.692	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	115.380%
35) Dibromofluoromethane	8.021	113	353037	48.373	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.740%
50) Toluene-d8	10.443	98	1510141	48.871	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.740%
62) 4-Bromofluorobenzene	12.733	95	449692	42.552	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.100%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	133313	28.793	ug/l	98
25) 2-Butanone	7.345	43	19751	3.433	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
Data File : VN068623.D
Acq On : 23 Sep 2021 16:35
Operator : JC/MD
Sample : M3826-05 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRIP-COMP-AUD-1181-1187

Quant Time: Sep 23 17:25:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

