

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
 Data File : VN068628.D
 Acq On : 23 Sep 2021 18:42
 Operator : JC/MD
 Sample : M3826-05 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRIP-COMP-AUD-1181-1187

Manual Integrations
 APPROVED

MMDadoda
 9/24/2021 6:59:01 PM

Quant Time: Sep 24 01:58:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	617392	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	1074762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	892521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	333002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	510422	53.056	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.120%			
35) Dibromofluoromethane	8.024	113	353439	49.109	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 98.220%			
50) Toluene-d8	10.443	98	1427160	46.821	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 93.640%			
62) 4-Bromofluorobenzene	12.733	95	442309	42.429	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 84.860%			
Target Compounds						
11) Tert butyl alcohol	5.366	59	10379m	6.996	ug/l	Qvalue
16) Acetone	4.288	43	806282	229.719	ug/l	99
25) 2-Butanone	7.340	43	126149	21.469	ug/l	99
95) Naphthalene	15.509	128	28278	3.304	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
Data File : VN068628.D
Acq On : 23 Sep 2021 18:42
Operator : JC/MD
Sample : M3826-05 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRIP-COMP-AUD-1181-1187

Manual Integrations
APPROVED

MMDadoda
9/24/2021 6:59:01 PM

Quant Time: Sep 24 01:58:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

