

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
 Data File : VN068631.D
 Acq On : 23 Sep 2021 19:58
 Operator : JC/MD
 Sample : M3871-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB02-BL

Quant Time: Sep 24 01:58:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	643424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1071364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	939066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	341243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	542908	54.175	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.340%	
35) Dibromofluoromethane	8.024	113	371151	51.781	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.560%	
50) Toluene-d8	10.446	98	1486432	48.920	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.840%	
62) 4-Bromofluorobenzene	12.733	95	479923	46.183	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.360%	
Target Compounds						
95) Naphthalene	15.509	128	6396	1.973	ug/l	Qvalue # 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092321\
Data File : VN068631.D
Acq On : 23 Sep 2021 19:58
Operator : JC/MD
Sample : M3871-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB02-BL

Quant Time: Sep 24 01:58:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

