

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084098.D
 Acq On : 23 Sep 2024 10:07
 Operator : JC\MD
 Sample : VN0923WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0923WBL01

Quant Time: Sep 24 01:27:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304776	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	258041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	133229	50.337	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.680%	
35) Dibromofluoromethane	8.165	113	99082	50.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	10.565	98	362159	53.691	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.380%	
62) 4-Bromofluorobenzene	12.847	95	129090	46.387	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084098.D
Acq On : 23 Sep 2024 10:07
Operator : JC\MD
Sample : VN0923WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0923WBL01

Quant Time: Sep 24 01:27:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

