

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084103.D
 Acq On : 23 Sep 2024 12:17
 Operator : JC\MD
 Sample : P4003-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20240912

Quant Time: Sep 24 01:30:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	238994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121687	49.723	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.440%
35) Dibromofluoromethane	8.171	113	90590	50.379	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.760%
50) Toluene-d8	10.565	98	338899	54.517	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.040%
62) 4-Bromofluorobenzene	12.847	95	121782	47.483	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.960%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	19923	10.561	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084103.D
Acq On : 23 Sep 2024 12:17
Operator : JC\MD
Sample : P4003-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20240912

Quant Time: Sep 24 01:30:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

