

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084110.D
 Acq On : 23 Sep 2024 15:29
 Operator : JC\MD
 Sample : P4003-12DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D2-20240912DL

Quant Time: Sep 24 01:33:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278606	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	228243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121249	51.334	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.660%	
35) Dibromofluoromethane	8.165	113	91689	51.407	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.820%	
50) Toluene-d8	10.565	98	323972	52.542	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	12.847	95	111797	43.946	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 87.900%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.347	130	52670	28.148	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084110.D
Acq On : 23 Sep 2024 15:29
Operator : JC\MD
Sample : P4003-12DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D2-20240912DL

Quant Time: Sep 24 01:33:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

