

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084112.D
 Acq On : 23 Sep 2024 16:17
 Operator : JC\MD
 Sample : P4003-10DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20240912DL

Quant Time: Sep 24 01:33:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281557	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125671	52.537	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.080%			
35) Dibromofluoromethane	8.165	113	91301	50.653	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.300%			
50) Toluene-d8	10.565	98	339900	54.547	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.100%			
62) 4-Bromofluorobenzene	12.847	95	115643	44.982	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.960%			
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	61546	32.547	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084112.D
Acq On : 23 Sep 2024 16:17
Operator : JC\MD
Sample : P4003-10DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20240912DL

Quant Time: Sep 24 01:33:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

