

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084115.D
 Acq On : 23 Sep 2024 17:29
 Operator : JC\MD
 Sample : P4003-16DL 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE108D2-20240913DL

Quant Time: Sep 24 04:59:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	149941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263543	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117558	51.996	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.000%
35) Dibromofluoromethane	8.171	113	85685	50.786	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.580%
50) Toluene-d8	10.565	98	317555	54.445	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.880%
62) 4-Bromofluorobenzene	12.847	95	108667	45.157	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.320%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	53364	30.149	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084115.D
Acq On : 23 Sep 2024 17:29
Operator : JC\MD
Sample : P4003-16DL 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE108D2-20240913DL

Quant Time: Sep 24 04:59:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

