

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084116.D
 Acq On : 23 Sep 2024 17:53
 Operator : JC\MD
 Sample : P4136-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Sep 24 04:59:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	150480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261972	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118067	52.034	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.060%	
35) Dibromofluoromethane	8.165	113	87214	52.003	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.000%	
50) Toluene-d8	10.565	98	317201	54.710	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.420%	
62) 4-Bromofluorobenzene	12.847	95	108313	45.280	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 90.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084116.D
Acq On : 23 Sep 2024 17:53
Operator : JC\MD
Sample : P4136-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Sep 24 04:59:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

