

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
 Data File : VN084119.D
 Acq On : 23 Sep 2024 19:05
 Operator : JC\MD
 Sample : P4136-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Sep 24 05:00:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	225678	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89361	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118904	51.035	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.060%
35) Dibromofluoromethane	8.165	113	87279	50.605	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	10.565	98	322780	54.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.280%
62) 4-Bromofluorobenzene	12.847	95	109204	44.393	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092324\
Data File : VN084119.D
Acq On : 23 Sep 2024 19:05
Operator : JC\MD
Sample : P4136-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Sep 24 05:00:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

