

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051455.D
 Acq On : 24 Sep 2018 12:41
 Operator : MD\SY
 Sample : J4770-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-092118-B

Quant Time: Sep 25 03:12:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	655099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	987261	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	838058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	320181	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	446730	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	7.59	113	393230	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	10.09	98	1401521	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
62) 4-Bromofluorobenzene	12.40	95	382136	39.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.46%	
Target Compounds						
18) Methyl Acetate	4.33	43	7850	1.256	ug/l	95
20) Methylene Chloride	4.55	84	2054171	252.961	ug/l	99
43) Isopropyl Acetate	8.17	43	214800	18.071	ug/l	99
95) Naphthalene	15.13	128	123499	16.749	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051455.D
Acq On : 24 Sep 2018 12:41
Operator : MD\SY
Sample : J4770-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-092118-B

Quant Time: Sep 25 03:12:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

