

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051458.D
 Acq On : 24 Sep 2018 13:55
 Operator : MD\SY
 Sample : J4770-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-092118-E

Quant Time: Sep 25 03:36:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	701258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1055466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	883971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	326259	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	463314	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
35) Dibromofluoromethane	7.59	113	405591	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
50) Toluene-d8	10.09	98	1496130	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
62) 4-Bromofluorobenzene	12.40	95	405994	39.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.98%	
Target Compounds						
18) Methyl Acetate	4.33	43	8623	1.289	ug/l	95
20) Methylene Chloride	4.55	84	432407	49.744	ug/l	100
43) Isopropyl Acetate	8.17	43	215994	16.997	ug/l	98
44) Trichloroethene	8.84	130	9728	1.021	ug/l	72
95) Naphthalene	15.13	128	82918	13.780	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051458.D
Acq On : 24 Sep 2018 13:55
Operator : MD\SY
Sample : J4770-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-092118-E

Quant Time: Sep 25 03:36:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

