

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051466.D
 Acq On : 24 Sep 2018 17:11
 Operator : MD\SY
 Sample : J5061-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230-KV-A

Quant Time: Sep 25 04:10:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	620918	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	940364	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	758610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	260505	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	425545	54.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.42%	
35) Dibromofluoromethane	7.59	113	328637	43.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.40%	
50) Toluene-d8	10.09	98	1311604	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
62) 4-Bromofluorobenzene	12.40	95	335186	36.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.18%	
Target Compounds						
18) Methyl Acetate	4.34	43	9392	1.585	ug/l #	74
20) Methylene Chloride	4.55	84	1287898	167.329	ug/l	99
95) Naphthalene	15.13	128	12115	9.094	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051466.D
Acq On : 24 Sep 2018 17:11
Operator : MD\SY
Sample : J5061-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230-KV-A

Quant Time: Sep 25 04:10:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

