

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051467.D
 Acq On : 24 Sep 2018 17:35
 Operator : MD\SY
 Sample : J5061-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230-KV-B

Quant Time: Sep 25 04:13:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	633996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	960767	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	791950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	305752	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	429186	54.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.08%	
35) Dibromofluoromethane	7.59	113	310179	40.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.74%	
50) Toluene-d8	10.09	98	1365552	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	12.40	95	366746	39.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.38%	
Target Compounds						
18) Methyl Acetate	4.34	43	8258	1.365	ug/l #	65
20) Methylene Chloride	4.55	84	182791	23.259	ug/l	99
30) Chloroform	7.37	83	17149	1.268	ug/l	97
68) m/p-Xylenes	11.62	106	21835	1.649	ug/l	98
95) Naphthalene	15.13	128	52973	11.954	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051467.D
Acq On : 24 Sep 2018 17:35
Operator : MD\SY
Sample : J5061-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230-KV-B

Quant Time: Sep 25 04:13:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

