

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051468.D
 Acq On : 24 Sep 2018 18:00
 Operator : MD\SY
 Sample : J5061-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230-KV-C

Quant Time: Sep 25 04:21:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	641154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	974509	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	801203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	295998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	438106	54.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.10%	
35) Dibromofluoromethane	7.59	113	332155	42.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.24%	
50) Toluene-d8	10.09	98	1372936	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
62) 4-Bromofluorobenzene	12.40	95	365377	38.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.98%	
Target Compounds						
18) Methyl Acetate	4.33	43	7156	1.170	ug/l #	75
20) Methylene Chloride	4.55	84	225382	28.358	ug/l	99
95) Naphthalene	15.13	128	18445	9.451	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051468.D
Acq On : 24 Sep 2018 18:00
Operator : MD\SY
Sample : J5061-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230-KV-C

Quant Time: Sep 25 04:21:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

