

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092418\
 Data File : VN051470.D
 Acq On : 24 Sep 2018 18:49
 Operator : MD\SY
 Sample : J4770-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-02-092118-C

Manual Integrations
 APPROVED

MMDadoda
 9/25/2018 11:31:04 AM

Quant Time: Sep 25 04:30:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	670988	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1018276	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	846572	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	304857	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	450305	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
35) Dibromofluoromethane	7.59	113	392824	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
50) Toluene-d8	10.09	98	1436273	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
62) 4-Bromofluorobenzene	12.40	95	381383	38.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.90%	
Target Compounds						
18) Methyl Acetate	4.34	43	8649	1.351	ug/l #	73
20) Methylene Chloride	4.55	84	289107	34.759	ug/l	99
30) Chloroform	7.37	83	19050m	1.331	ug/l	
43) Isopropyl Acetate	8.17	43	199126	16.242	ug/l	97
95) Naphthalene	15.13	128	16479	9.264	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092418\
Data File : VN051470.D
Acq On : 24 Sep 2018 18:49
Operator : MD\SY
Sample : J4770-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-02-092118-C

Manual Integrations
APPROVED

MMDadoda
9/25/2018 11:31:04 AM

Quant Time: Sep 25 04:30:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

