

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092420\
 Data File : VN063620.D
 Acq On : 24 Sep 2020 11:48
 Operator : JC/MD
 Sample : VN0924MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0924MBL01

Quant Time: Sep 24 16:56:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	210170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	395617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	385328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	154937	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	178198	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
35) Dibromofluoromethane	7.55	113	131504	48.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.04%	
50) Toluene-d8	10.06	98	502865	48.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.66%	
62) 4-Bromofluorobenzene	12.38	95	184916	48.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092420\
Data File : VN063620.D
Acq On : 24 Sep 2020 11:48
Operator : JC/MD
Sample : VN0924MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0924MBL01

Quant Time: Sep 24 16:56:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

