

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063621.D
 Acq On : 24 Sep 2020 12:12
 Operator : JC/MD
 Sample : VN0924WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0924WBL01

Quant Time: Sep 24 16:57:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	212611	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	392850	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	396689	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165500	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	177713	49.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.20%	
35) Dibromofluoromethane	7.55	113	131375	48.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.62%	
50) Toluene-d8	10.06	98	508807	49.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	196866	52.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.02%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063621.D
Acq On : 24 Sep 2020 12:12
Operator : JC/MD
Sample : VN0924WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0924WBL01

Quant Time: Sep 24 16:57:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

