

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063627.D
 Acq On : 24 Sep 2020 14:40
 Operator : JC/MD
 Sample : L4083-01DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20200916DL

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:16:12 PM

Quant Time: Sep 25 04:03:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	199853	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	377734	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	382265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161426	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	169604	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
35) Dibromofluoromethane	7.55	113	124438	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
50) Toluene-d8	10.06	98	488200	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.38	95	190544	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	4102	1.842	ug/l #	91
12) 1,1-Dichloroethene	3.71	96	2521m	1.118	ug/l	
44) Trichloroethene	8.80	130	175927	63.140	ug/l	99
64) Tetrachloroethene	10.60	164	899	0.345	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063627.D
Acq On : 24 Sep 2020 14:40
Operator : JC/MD
Sample : L4083-01DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20200916DL

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:16:12 PM

Quant Time: Sep 25 04:03:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

