

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063630.D
 Acq On : 24 Sep 2020 15:52
 Operator : JC/MD
 Sample : L4119-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D2-20200921

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:16:16 PM

Quant Time: Sep 25 16:34:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	195922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	368114	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	377407	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	160218	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	169045	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
35) Dibromofluoromethane	7.55	113	121051	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
50) Toluene-d8	10.06	98	487046	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	12.38	95	188403	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.73	101	7008m	3.210	ug/l	Qvalue
27) cis-1,2-Dichloroethene	6.79	96	1151	0.415	ug/l	16
30) Chloroform	7.33	83	3017	0.618	ug/l	86
44) Trichloroethene	8.80	130	1457962	536.938	ug/l	100
64) Tetrachloroethene	10.61	164	1676	0.652	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063630.D
Acq On : 24 Sep 2020 15:52
Operator : JC/MD
Sample : L4119-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D2-20200921

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:16:16 PM

Quant Time: Sep 25 16:34:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

