

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063643.D
 Acq On : 24 Sep 2020 21:04
 Operator : JC/MD
 Sample : L4119-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE126D3-20200921

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:17:04 PM

Quant Time: Sep 25 05:16:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	194356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	365229	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	370175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	156970	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	171822	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
35) Dibromofluoromethane	7.55	113	124800	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	10.06	98	474235	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
62) 4-Bromofluorobenzene	12.38	95	181668	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
Target Compounds						
3) Chloromethane	2.04	50	1393	0.478	ug/l	94
16) Acetone	3.78	43	3773	3.548	ug/l	96
44) Trichloroethene	8.81	130	6364m	2.362	ug/l	
64) Tetrachloroethene	10.60	164	3544	1.406	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063643.D
Acq On : 24 Sep 2020 21:04
Operator : JC/MD
Sample : L4119-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE126D3-20200921

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:17:04 PM

Quant Time: Sep 25 05:16:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

