

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063652.D
 Acq On : 25 Sep 2020 1:50
 Operator : JC/MD
 Sample : L3871-29
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-092320

Quant Time: Sep 25 06:04:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	197061	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	369384	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	377399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	158439	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	170404	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	7.55	113	120839	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	10.06	98	482278	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.38	95	186253	52.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.68%	
Target Compounds						
16) Acetone	3.78	43	7578	7.029	ug/l	95
20) Methylene Chloride	4.51	84	65741	24.068	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063652.D
Acq On : 25 Sep 2020 1:50
Operator : JC/MD
Sample : L3871-29
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-092320

Quant Time: Sep 25 06:04:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

