

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063654.D
 Acq On : 25 Sep 2020 2:37
 Operator : JC/MD
 Sample : PB131850ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131850ZHE#16

Quant Time: Sep 25 06:07:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	203237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	396328	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	413487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	173114	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	176457	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
35) Dibromofluoromethane	7.55	113	129498	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	10.06	98	507177	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
62) 4-Bromofluorobenzene	12.38	95	199494	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
16) Acetone	3.78	43	11553	10.391	ug/l	94
18) Methyl Acetate	4.28	43	4276	1.515	ug/l #	64
20) Methylene Chloride	4.51	84	9167	3.254	ug/l	90
43) Isopropyl Acetate	8.13	43	72014	9.719	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063654.D
Acq On : 25 Sep 2020 2:37
Operator : JC/MD
Sample : PB131850ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131850ZHE#16

Quant Time: Sep 25 06:07:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

