

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063655.D
 Acq On : 25 Sep 2020 3:01
 Operator : JC/MD
 Sample : PB131852TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131852TB

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:15:49 PM

Quant Time: Sep 25 06:09:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	218101	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	422529	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	437013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183709	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	190802	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	7.55	113	138009	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
50) Toluene-d8	10.06	98	548915	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
62) 4-Bromofluorobenzene	12.38	95	213324	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
Target Compounds						
16) Acetone	3.78	43	12778m	10.709	ug/l	
18) Methyl Acetate	4.28	43	4897m	1.617	ug/l	
20) Methylene Chloride	4.51	84	17918	5.927	ug/l	89
43) Isopropyl Acetate	8.13	43	74590	9.442	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063655.D
Acq On : 25 Sep 2020 3:01
Operator : JC/MD
Sample : PB131852TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131852TB

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:15:49 PM

Quant Time: Sep 25 06:09:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

