

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063656.D
 Acq On : 25 Sep 2020 3:25
 Operator : JC/MD
 Sample : PB131852ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131852ZHE#01

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:15:36 PM

9/28/2020 5:15:36 PM

Quant Time: Sep 25 06:12:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	403146	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	407752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	174327	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	180096	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
35) Dibromofluoromethane	7.55	113	132968	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	10.06	98	518208	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
62) 4-Bromofluorobenzene	12.38	95	203854	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
Target Compounds						
16) Acetone	3.79	43	13609	12.185	ug/l #	84
18) Methyl Acetate	4.29	43	5058m	1.784	ug/l	
20) Methylene Chloride	4.51	84	18065	6.384	ug/l #	84
43) Isopropyl Acetate	8.13	43	74964	9.946	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063656.D
Acq On : 25 Sep 2020 3:25
Operator : JC/MD
Sample : PB131852ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131852ZHE#01

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:15:36 PM

Quant Time: Sep 25 06:12:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

