

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063659.D
 Acq On : 25 Sep 2020 4:37
 Operator : JC/MD
 Sample : L3870-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-092220

Quant Time: Sep 25 06:17:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	199980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	393481	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	403078	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	163356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	177346	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
35) Dibromofluoromethane	7.55	113	126957	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
50) Toluene-d8	10.06	98	511005	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
62) 4-Bromofluorobenzene	12.38	95	192014	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
Target Compounds						
16) Acetone	3.78	43	26193	23.941	ug/l	99
18) Methyl Acetate	4.29	43	5571	2.006	ug/l #	66
20) Methylene Chloride	4.51	84	14490	5.227	ug/l	88
43) Isopropyl Acetate	8.13	43	49497	6.728	ug/l	98
95) Naphthalene	15.10	128	24373	2.335	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063659.D
Acq On : 25 Sep 2020 4:37
Operator : JC/MD
Sample : L3870-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-092220

Quant Time: Sep 25 06:17:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

