

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063665.D
 Acq On : 25 Sep 2020 7:01
 Operator : JC/MD
 Sample : L4119-18 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D2-20200922

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 5:15:17 PM

Quant Time: Sep 25 07:43:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	187382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	360482	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	358946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	152008	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	168898	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
35) Dibromofluoromethane	7.55	113	122715	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
50) Toluene-d8	10.06	98	466991	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	12.38	95	179380	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	3999m	1.916	ug/l	
12) 1,1-Dichloroethene	3.70	96	677	0.320	ug/l #	66
44) Trichloroethene	8.80	130	208979	78.592	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063665.D
Acq On : 25 Sep 2020 7:01
Operator : JC/MD
Sample : L4119-18 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D2-20200922

Manual Integrations
APPROVED

MMDadoda
9/28/2020 5:15:17 PM

Quant Time: Sep 25 07:43:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

