

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063668.D
 Acq On : 25 Sep 2020 8:13
 Operator : JC/MD
 Sample : L4119-05 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D2-20200921

Quant Time: Sep 25 08:40:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	187464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	361165	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	354861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	149198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	166831	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	7.55	113	118473	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	10.06	98	458765	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.38	95	176752	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
Target Compounds						
44) Trichloroethene	8.80	130	285129	107.028	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063668.D
Acq On : 25 Sep 2020 8:13
Operator : JC/MD
Sample : L4119-05 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D2-20200921

Quant Time: Sep 25 08:40:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

