

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092420\
 Data File : VN063669.D
 Acq On : 25 Sep 2020 8:37
 Operator : JC/MD
 Sample : L4098-19 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRIP-LIQUID

Quant Time: Sep 25 09:03:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	191969	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	352805	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	360971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	163278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	163559	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	7.55	113	118127	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
50) Toluene-d8	10.06	98	456157	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	12.38	95	184322	54.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.50%	
Target Compounds						
16) Acetone	3.78	43	73611	70.091	ug/l	99
20) Methylene Chloride	4.49	84	3330	1.251	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092420\
Data File : VN063669.D
Acq On : 25 Sep 2020 8:37
Operator : JC/MD
Sample : L4098-19 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRIP-LIQUID

Quant Time: Sep 25 09:03:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

