

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN092421\  
 Data File : VN068647.D  
 Acq On : 24 Sep 2021 18:42  
 Operator : JC/MD  
 Sample : PB139263TB  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB139263TB

Quant Time: Sep 25 01:51:50 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092321W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 23 13:52:37 2021  
 Response via : Initial Calibration


| Compound                      | R.T.           | QIon | Response   | Conc      | Units  | Dev(Min) |
|-------------------------------|----------------|------|------------|-----------|--------|----------|
| -----                         |                |      |            |           |        |          |
| Internal Standards            |                |      |            |           |        |          |
| 1) Pentafluorobenzene         | 8.088          | 168  | 533748     | 50.000    | ug/l   | # 0.00   |
| 34) 1,4-Difluorobenzene       | 8.971          | 114  | 886513     | 50.000    | ug/l   | 0.00     |
| 63) Chlorobenzene-d5          | 11.747         | 117  | 861572     | 50.000    | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4    | 13.678         | 152  | 282079     | 50.000    | ug/l   | 0.00     |
|                               |                |      |            |           |        |          |
| System Monitoring Compounds   |                |      |            |           |        |          |
| 33) 1,2-Dichloroethane-d4     | 8.440          | 65   | 404801     | 48.568    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 61 - 141 |      | Recovery = | 97.140%   |        |          |
| 35) Dibromofluoromethane      | 8.024          | 113  | 305068     | 51.430    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 69 - 133 |      | Recovery = | 102.860%  |        |          |
| 50) Toluene-d8                | 10.443         | 98   | 1251671    | 49.784    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 65 - 126 |      | Recovery = | 99.560%   |        |          |
| 62) 4-Bromofluorobenzene      | 12.734         | 95   | 403734     | 46.953    | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 58 - 135 |      | Recovery = | 93.900%   |        |          |
|                               |                |      |            |           |        |          |
| Target Compounds              |                |      |            |           |        |          |
|                               |                |      |            |           | Qvalue |          |
| 5) Bromomethane               | 2.883          | 94   | 987        | Below Cal | #      | 71       |
| 6) Chloroethane               | 3.025          | 64   | 185        | Below Cal | #      | 44       |
| 13) Acrolein                  | 4.044          | 56   | 126        | 7.129     | ug/l # | 14       |
| 16) Acetone                   | 4.293          | 43   | 101905     | 23.317    | ug/l   | 96       |
| 20) Methylene Chloride        | 5.106          | 84   | 28659      | 2.657     | ug/l   | 93       |
| 25) 2-Butanone                | 7.337          | 43   | 24943      | 4.910     | ug/l   | 94       |
| 29) Tetrahydrofuran           | 7.691          | 42   | 876        | 0.271     | ug/l   | 93       |
| 31) Cyclohexane               | 8.086          | 56   | 12586      | 1.076     | ug/l # | 1        |
| 36) 1,1-Dichloropropene       | 8.091          | 75   | 44384      | 4.957     | ug/l # | 52       |
| 37) Ethyl Acetate             | 7.337          | 43   | 24943      | 2.454     | ug/l # | 69       |
| 38) Carbon Tetrachloride      | 8.091          | 117  | 55355      | 5.986     | ug/l # | 17       |
| 41) Methacrylonitrile         | 7.699          | 41   | 1143       | 0.213     | ug/l # | 33       |
| 43) Isopropyl Acetate         | 8.563          | 43   | 52766      | 3.034     | ug/l # | 91       |
| 45) 1,2-Dichloropropane       | 9.628          | 63   | 3370       | 0.487     | ug/l # | 44       |
| 48) Methyl methacrylate       | 9.545          | 41   | 13221      | 1.665     | ug/l # | 12       |
| 49) 1,4-Dioxane               | 9.614          | 88   | 188        | 1.623     | ug/l # | 1        |
| 51) 4-Methyl-2-Pentanone      | 10.333         | 43   | 14366      | 1.366     | ug/l   | 100      |
| 54) cis-1,3-Dichloropropene   | 10.245         | 75   | 77904      | 7.126     | ug/l # | 64       |
| 56) Ethyl methacrylate        | 10.870         | 69   | 432        | 0.742     | ug/l # | 1        |
| 57) 1,3-Dichloropropane       | 11.092         | 76   | 49671      | 4.407     | ug/l # | 43       |
| 59) 2-Hexanone                | 11.092         | 43   | 810674     | 107.202   | ug/l # | 52       |
| 76) 1,2,3-Trichloropropane    | 12.734         | 75   | 232539     | 33.341    | ug/l # | 100      |
| 80) 1,3,5-Trimethylbenzene    | 13.063         | 105  | 4840       | 0.244     | ug/l   | 94       |
| 81) trans-1,4-Dichloro-2-b... | 12.734         | 75   | 232539     | 94.179    | ug/l # | 6        |
| 95) Naphthalene               | 15.507         | 128  | 2469       | 1.773     | ug/l # | 83       |
| -----                         |                |      |            |           |        |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN092421\  
Data File : VN068647.D  
Acq On : 24 Sep 2021 18:42  
Operator : JC/MD  
Sample : PB139263TB  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB

Quant Time: Sep 25 01:51:50 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N092321W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 23 13:52:37 2021  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.088 min Scan# 2274

Delta R.T. 0.000 min





Lab File: VN068647.D

Acq: 24 Sep 2021 18:42

Tgt Ion:168 Resp: 533748

Ion Ratio Lower Upper

168 100  
99 59.5 62.9 94.3#



#5

Bromomethane

Concen: Below Cal

RT: 2.883 min Scan# 333

Delta R.T. 0.046 min




Lab File: VN068647.D


Acq: 24 Sep 2021 18:42

Tgt Ion: 94 Resp: 987

Ion Ratio Lower Upper

94 100  
96 64.4 73.7 110.5#









#6  
Chloroethane  
Concen: Below Cal  
RT: 3.025 min Scan# 386  
Delta R.T. 0.016 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion: 64 Resp: 185  
Ion Ratio Lower Upper  
64 100  
66 0.0 24.7 37.1#



#13  
Acrolein  
Concen: 7.129 ug/l  
RT: 4.044 min Scan# 766  
Delta R.T. -0.000 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 56 Resp: 126  
Ion Ratio Lower Upper  
56 100  
55 0.0 57.4 86.0#









#16  
Acetone  
Concen: 23.317 ug/l  
RT: 4.293 min Scan# 859  
Delta R.T. 0.005 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion: 43 Resp: 101905  
Ion Ratio Lower Upper  
43 100  
58 28.7 21.2 31.8



#20  
Methylene Chloride  
Concen: 2.657 ug/l  
RT: 5.106 min Scan# 1162  
Delta R.T. 0.008 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 84 Resp: 28659  
Ion Ratio Lower Upper  
84 100  
49 127.6 110.9 166.3  
51 34.8 32.6 49.0  
86 62.6 51.6 77.4









#25  
2-Butanone  
Concen: 4.910 ug/l  
RT: 7.337 min Scan# 1994  
Delta R.T. 0.002 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampled :  
PB139263TB


Tgt Ion: 43 Resp: 24943  
Ion Ratio Lower Upper  
43 100  
72 21.6 19.7 29.5



#29  
Tetrahydrofuran  
Concen: 0.271 ug/l  
RT: 7.691 min Scan# 2126  
Delta R.T. 0.008 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 42 Resp: 876  
Ion Ratio Lower Upper  
42 100  
72 36.1 32.8 49.2  
71 43.0 31.2 46.8





#31

Cyclohexane

Concen: 1.076 ug/l

RT: 8.086 min Scan# 2273

Delta R.T. -0.010 min

Lab File: VN068647.D

Acq: 24 Sep 2021 18:42

Tgt Ion: 56 Resp: 12586

Ion Ratio Lower Upper

56 100

69 181.0 27.3 40.9#





84 113.1 64.9 97.3#

Instrument :

MSVOA\_N

ClientSampleId :

PB139263TB



#33

1,2-Dichloroethane-d4

Concen: 48.568 ug/l

RT: 8.440 min Scan# 2405

Delta R.T. -0.000 min

Lab File: VN068647.D




Acq: 24 Sep 2021 18:42


Tgt Ion: 65 Resp: 404801

Ion Ratio Lower Upper

65 100

67 51.7 0.0 98.6









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.971 min Scan# 2603  
Delta R.T. 0.003 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB

Tgt Ion:114 Resp: 886513  
Ion Ratio Lower Upper  
114 100  
63 22.2 0.0 49.0  
88 15.7 0.0 36.6




#35  
Dibromofluoromethane  
Concen: 51.430 ug/l  
RT: 8.024 min Scan# 2250  
Delta R.T. 0.000 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion:113 Resp: 305068  
Ion Ratio Lower Upper  
113 100  
111 103.3 83.8 125.8  
192 17.3 13.4 20.0











#36  
1,1-Dichloropropene  
Concen: 4.957 ug/l  
RT: 8.091 min Scan# 2275  
Delta R.T. -0.132 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion: 75 Resp: 44384  
Ion Ratio Lower Upper  
75 100  
110 8.8 15.7 47.0#  
77 0.0 24.6 37.0#



#37  
Ethyl Acetate  
Concen: 2.454 ug/l  
RT: 7.337 min Scan# 1994  
Delta R.T. -0.076 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 43 Resp: 24943  
Ion Ratio Lower Upper  
43 100  
61 0.0 10.6 16.0#  
70 0.0 8.1 12.1#









#38  
Carbon Tetrachloride  
Concen: 5.986 ug/l  
RT: 8.091 min Scan# 2275  
Delta R.T. -0.118 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion:117 Resp: 55355  
Ion Ratio Lower Upper  
117 100  
119 5.5 75.4 113.2#  
121 0.0 23.5 35.3#



#41  
Methacrylonitrile  
Concen: 0.213 ug/l  
RT: 7.699 min Scan# 2129  
Delta R.T. 0.061 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 41 Resp: 1143  
Ion Ratio Lower Upper  
41 100  
39 22.1 47.6 71.4#  
67 0.0 57.7 86.5#  
52 0.0 26.2 39.4#









#43  
Isopropyl Acetate  
Concen: 3.034 ug/l  
RT: 8.563 min Scan# 2451  
Delta R.T. 0.003 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampled :  
PB139263TB


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 43      | 100   |       |       |
| 61      | 17.3  | 19.0  | 28.4# |
| 87      | 11.5  | 9.5   | 14.3  |



#45  
1,2-Dichloropropane  
Concen: 0.487 ug/l  
RT: 9.628 min Scan# 2848  
Delta R.T. 0.137 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 63      | 100   |       |       |
| 65      | 0.0   | 24.3  | 36.5# |









#48  
Methyl methacrylate  
Concen: 1.665 ug/l  
RT: 9.545 min Scan# 2817  
Delta R.T. -0.016 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampled :  
PB139263TB


Tgt Ion: 41 Resp: 13221  
Ion Ratio Lower Upper  
41 100  
69 0.0 64.6 96.8#  
39 0.0 53.3 79.9#



#49  
1,4-Dioxane  
Concen: 1.623 ug/l  
RT: 9.614 min Scan# 2843  
Delta R.T. 0.045 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 88 Resp: 188  
Ion Ratio Lower Upper  
88 100  
43 1559319.1 28.8 43.2#  
58 6755.9 60.4 90.6#









#50  
Toluene-d8  
Concen: 49.784 ug/l  
RT: 10.443 min Scan# 3152  
Delta R.T. 0.000 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampled :  
PB139263TB


Tgt Ion: 98 Resp: 1251671  
Ion Ratio Lower Upper  
98 100  
100 63.9 51.7 77.5



#51  
4-Methyl-2-Pentanone  
Concen: 1.366 ug/l  
RT: 10.333 min Scan# 3111  
Delta R.T. 0.002 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 43 Resp: 14366  
Ion Ratio Lower Upper  
43 100  
58 36.0 28.6 43.0









#54  
 cis-1,3-Dichloropropene  
 Concen: 7.126 ug/l  
 RT: 10.245 min Scan# 3078  
 Delta R.T. 0.054 min  
 Lab File: VN068647.D  
 Acq: 24 Sep 2021 18:42




Instrument :  
 MSVOA\_N  
 ClientSampled :  
 PB139263TB


Tgt Ion: 75 Resp: 77904  
 Ion Ratio Lower Upper  
 75 100  
 77 0.6 24.6 37.0#  
 39 42.5 50.7 76.1#



#56  
 Ethyl methacrylate  
 Concen: 0.742 ug/l  
 RT: 10.870 min Scan# 3311  
 Delta R.T. 0.108 min  
 Lab File: VN068647.D  
 Acq: 24 Sep 2021 18:42

Tgt Ion: 69 Resp: 432  
 Ion Ratio Lower Upper  
 69 100  
 41 7107.9 58.6 88.0#  
 39 3447.9 38.3 57.5#









#57  
1,3-Dichloropropane  
Concen: 4.407 ug/l  
RT: 11.092 min Scan# 3394  
Delta R.T. 0.045 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion: 76 Resp: 49671  
Ion Ratio Lower Upper  
76 100  
78 0.0 25.7 38.5#



#59  
2-Hexanone  
Concen: 107.202 ug/l  
RT: 11.092 min Scan# 3394  
Delta R.T. 0.005 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 43 Resp: 810674  
Ion Ratio Lower Upper  
43 100  
58 16.6 24.7 74.1#









#62  
4-Bromofluorobenzene  
Concen: 46.953 ug/l  
RT: 12.734 min Scan# 4006  
Delta R.T. 0.003 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB

Tgt Ion: 95 Resp: 403734  
Ion Ratio Lower Upper  
95 100  
174 75.2 0.0 122.0  
176 72.7 0.0 117.6




#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.747 min Scan# 3638  
Delta R.T. -0.000 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 117 Resp: 861572  
Ion Ratio Lower Upper  
117 100  
82 57.9 53.1 79.7  
119 32.4 25.5 38.3











#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.678 min Scan# 4358  
Delta R.T. 0.003 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampled :  
PB139263TB


Tgt Ion:152 Resp: 282079  
Ion Ratio Lower Upper  
152 100  
115 60.7 36.1 108.5  
150 158.8 0.0 351.8



#76  
1,2,3-Trichloropropane  
Concen: 33.341 ug/l  
RT: 12.734 min Scan# 4006  
Delta R.T. -0.147 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Tgt Ion: 75 Resp: 232539  
Ion Ratio Lower Upper  
75 100  
77 1.0 0.0 0.0#









#80  
1,3,5-Trimethylbenzene  
Concen: 0.244 ug/l  
RT: 13.063 min Scan# 4129  
Delta R.T. 0.002 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42




Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB


Tgt Ion:105 Resp: 4840  
Ion Ratio Lower Upper  
105 100  
120 48.0 22.2 66.6



#81  
trans-1,4-Dichloro-2-butene  
Concen: 94.179 ug/l  
RT: 12.734 min Scan# 4006  
Delta R.T. 0.108 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42


Tgt Ion: 75 Resp: 232539  
Ion Ratio Lower Upper  
75 100  
53 0.1 87.6 131.4#  
89 0.0 37.8 56.8#





#95  
Naphthalene  
Concen: 1.773 ug/l  
RT: 15.507 min Scan# 5040  
Delta R.T. -0.000 min  
Lab File: VN068647.D  
Acq: 24 Sep 2021 18:42

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB139263TB



Tgt Ion:128 Resp: 2469  
Ion Ratio Lower Upper  
128 100  
127 6.2 10.6 15.8#  
129 4.7 8.6 12.8#

