

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068636.D
 Acq On : 24 Sep 2021 13:21
 Operator : JC/MD
 Sample : VN0924MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0924MBL01

Quant Time: Sep 24 17:18:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	543959	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	885684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	720733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	264540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	411139	48.398	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.800%
35) Dibromofluoromethane	8.021	113	295858	49.898	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.800%
50) Toluene-d8	10.443	98	1166000	46.420	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.840%
62) 4-Bromofluorobenzene	12.733	95	331260	38.560	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	77.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068636.D
Acq On : 24 Sep 2021 13:21
Operator : JC/MD
Sample : VN0924MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0924MBL01

Quant Time: Sep 24 17:18:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

