

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068644.D
 Acq On : 24 Sep 2021 17:26
 Operator : JC/MD
 Sample : M3911-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-CB-02-(0.2)

Quant Time: Sep 24 17:49:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	471786	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	769988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	669932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	242492	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	360807	48.986	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.980%
35) Dibromofluoromethane	8.024	113	275688	53.547	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.100%
50) Toluene-d8	10.443	98	1076005	49.273	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.540%
62) 4-Bromofluorobenzene	12.736	95	327134	43.802	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068644.D
Acq On : 24 Sep 2021 17:26
Operator : JC/MD
Sample : M3911-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-CB-02-(0.2)

Quant Time: Sep 24 17:49:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

