

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068648.D
 Acq On : 24 Sep 2021 19:08
 Operator : JC/MD
 Sample : M3823-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DKM-ENV-IC

Quant Time: Sep 25 01:52:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	502906	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	822296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	771256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	274501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	379506	48.319	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.640%
35) Dibromofluoromethane	8.024	113	286485	52.080	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.160%
50) Toluene-d8	10.443	98	1428005	61.233	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	122.460%
62) 4-Bromofluorobenzene	12.733	95	342615	42.957	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.920%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	105050	26.642	ug/l	99
20) Methylene Chloride	5.103	84	23400	2.026	ug/l	91
25) 2-Butanone	7.337	43	25184	5.262	ug/l	95
43) Isopropyl Acetate	8.563	43	51817	3.213	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068648.D
Acq On : 24 Sep 2021 19:08
Operator : JC/MD
Sample : M3823-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DKM-ENV-IC

Quant Time: Sep 25 01:52:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

