

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068649.D
 Acq On : 24 Sep 2021 19:33
 Operator : JC/MD
 Sample : M3823-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DKM-ENV-IG

Quant Time: Sep 25 01:52:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	459297	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	768697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	709304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	255143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	362926	50.654	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.300%
35) Dibromofluoromethane	8.024	113	271319	52.774	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.540%
50) Toluene-d8	10.443	98	1112726	51.040	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.080%
62) 4-Bromofluorobenzene	12.734	95	338823	45.443	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.880%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	102785	29.400	ug/l	96
20) Methylene Chloride	5.093	84	23189	2.375	ug/l	95
25) 2-Butanone	7.340	43	23377	5.348	ug/l	94
43) Isopropyl Acetate	8.563	43	53305	3.535	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068649.D
Acq On : 24 Sep 2021 19:33
Operator : JC/MD
Sample : M3823-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DKM-ENV-IG

Quant Time: Sep 25 01:52:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

