

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068651.D
 Acq On : 24 Sep 2021 20:24
 Operator : JC/MD
 Sample : M3824-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-C

Manual Integrations
 APPROVED

MMDadoda
 9/25/2021 10:23:45 AM

Quant Time: Sep 25 01:52:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	461510	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	775099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	712447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	264652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	362234	50.307	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	100.620%	
35) Dibromofluoromethane	8.024	113	273140	52.688	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	105.380%	
50) Toluene-d8	10.446	98	1098581	49.975	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	99.960%	
62) 4-Bromofluorobenzene	12.733	95	335144	44.579	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	89.160%	
Target Compounds						Qvalue
16) Acetone	4.291	43	103094	29.326	ug/l	93
20) Methylene Chloride	5.090	84	23519m	2.416	ug/l	
25) 2-Butanone	7.345	43	24244	5.520	ug/l	93
43) Isopropyl Acetate	8.563	43	56294	3.703	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068651.D
Acq On : 24 Sep 2021 20:24
Operator : JC/MD
Sample : M3824-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-C

Manual Integrations
APPROVED

MMDadoda
9/25/2021 10:23:45 AM

Quant Time: Sep 25 01:52:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

