

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068654.D
 Acq On : 24 Sep 2021 21:40
 Operator : JC/MD
 Sample : M3773-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1508

Quant Time: Sep 25 01:53:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	492801	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	817601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	724245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	269821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	380490	49.467	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.940%
35) Dibromofluoromethane	8.024	113	290218	53.078	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.160%
50) Toluene-d8	10.443	98	1127033	48.604	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.734	95	341511	43.064	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.120%
Target Compounds						
						Qvalue
16) Acetone	4.288	43	291930	97.632	ug/l	95
20) Methylene Chloride	5.098	84	31222	3.508	ug/l	93
43) Isopropyl Acetate	8.566	43	60558	3.776	ug/l	# 76
52) Toluene	10.507	92	30196	1.966	ug/l	96
95) Naphthalene	15.507	128	8939	2.263	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068654.D
Acq On : 24 Sep 2021 21:40
Operator : JC/MD
Sample : M3773-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1508

Quant Time: Sep 25 01:53:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

