

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
 Data File : VN068655.D
 Acq On : 24 Sep 2021 22:05
 Operator : JC/MD
 Sample : M3762-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11934

Quant Time: Sep 25 01:53:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 13:52:37 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	488840	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	821294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	736957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	269064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	386837	50.730	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	101.460%	
35) Dibromofluoromethane	8.024	113	283843	51.656	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	103.320%	
50) Toluene-d8	10.443	98	1151977	49.457	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	98.920%	
62) 4-Bromofluorobenzene	12.734	95	349729	43.902	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	87.800%	
Target Compounds						
						Qvalue
16) Acetone	4.291	43	131859	37.906	ug/l	95
20) Methylene Chloride	5.103	84	31917	3.678	ug/l	92
43) Isopropyl Acetate	8.563	43	64921	4.030	ug/l	# 84
95) Naphthalene	15.509	128	8714	2.248	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092421\
Data File : VN068655.D
Acq On : 24 Sep 2021 22:05
Operator : JC/MD
Sample : M3762-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11934

Quant Time: Sep 25 01:53:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 13:52:37 2021
Response via : Initial Calibration

