

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084124.D
 Acq On : 24 Sep 2024 10:57
 Operator : JC\MD
 Sample : VN0924WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0924WBL01

Quant Time: Sep 25 01:29:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288254	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98488	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	127055	50.387	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.780%
35) Dibromofluoromethane	8.165	113	94453	51.184	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	10.565	98	341384	53.512	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.020%
62) 4-Bromofluorobenzene	12.847	95	119557	45.424	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084124.D
Acq On : 24 Sep 2024 10:57
Operator : JC\MD
Sample : VN0924WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0924WBL01

Quant Time: Sep 25 01:29:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

