

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084131.D
 Acq On : 24 Sep 2024 13:55
 Operator : JC\MD
 Sample : P4116-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUPE04-20240916

Quant Time: Sep 25 01:33:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	146558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	258216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	216916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	83851	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114164	51.661	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.320%
35) Dibromofluoromethane	8.165	113	86204	52.148	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.300%
50) Toluene-d8	10.565	98	307938	53.885	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.760%
62) 4-Bromofluorobenzene	12.847	95	106085	44.994	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084131.D
Acq On : 24 Sep 2024 13:55
Operator : JC\MD
Sample : P4116-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUPE04-20240916

Quant Time: Sep 25 01:33:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

