

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084133.D
 Acq On : 24 Sep 2024 14:43
 Operator : JC\MD
 Sample : P4116-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D3-20240916

Quant Time: Sep 25 01:34:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	91202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121176	51.908	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.820%	
35) Dibromofluoromethane	8.165	113	89493	51.815	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.620%	
50) Toluene-d8	10.565	98	322567	54.023	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.040%	
62) 4-Bromofluorobenzene	12.847	95	109090	44.283	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 88.560%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	21632	11.938	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084133.D
Acq On : 24 Sep 2024 14:43
Operator : JC\MD
Sample : P4116-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D3-20240916

Quant Time: Sep 25 01:34:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

