

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
 Data File : VN084135.D
 Acq On : 24 Sep 2024 15:31
 Operator : JC\MD
 Sample : P4116-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT191D1-20240917

Quant Time: Sep 25 01:35:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118323	51.039	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.080%	
35) Dibromofluoromethane	8.165	113	88595	51.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.920%	
50) Toluene-d8	10.565	98	323802	54.401	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.800%	
62) 4-Bromofluorobenzene	12.847	95	110412	44.962	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 89.920%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.365	101	4634	2.632	ug/l	90
12) 1,1-Dichloroethene	4.347	96	1086	0.653	ug/l	96
30) Chloroform	7.965	83	2112	0.566	ug/l	94
44) Trichloroethene	9.353	130	33494	18.543	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092424\
Data File : VN084135.D
Acq On : 24 Sep 2024 15:31
Operator : JC\MD
Sample : P4116-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT191D1-20240917

Quant Time: Sep 25 01:35:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

